

**CORRIGENDUM FOR MODIFICATION AND CLARIFICATION FOR  
SPECIFICATION**

Ref: NITJSR/MED/84/2019/S&P/221

Date: 04/03/2020

**Name of Equipment: EQUIPMENT PROCUREMENT OF ZETA POTENTIAL  
ANALYZER**

**Tender Reference No.:** NITJSR/MED/84/2019/S&P/184 Dated: 17/02/2020

**CPP ID No.:** 2020\_NITJS\_545168\_1

**Submitted Technical Specifications and Corrections**

<b>Parameter</b>	<b>Earlier</b>	<b>Modified</b>
<b>Zeta potential range:</b>	±1000 mV or higher	±500 mV or higher
<b>Particle size for measurement:</b>	0.5 nm to 50 µm	0.5 nm to 10 µm
<b>Refractive Index and transmittance:</b>	integrated	Optional
<b>Measuring Angle:</b>	15°	13°-15°, multiple